


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130121		
Name:	<u>Kyle Carrington</u>	ID#	<u>5336</u>	Date:	<u>6/8/08</u>
Agency:	<u>Fairbanks Airport Police</u>	Phone #:	<u>(907) 474-2555</u>		
Instrument Location:	<u>Fairbanks Airport Police</u> ✓				
Alco S/N:	<u>78895</u> ✓	Target Value:	<u>.081</u> ✓	High Pressure:	<u>1150 PSI</u>
Alco Test Value Average	<u>.080</u> ✓		<u>.080</u> ✓		
	1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco		
Signature					<u>0013</u> <u>6/24/08</u>
		(OVER)			


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

6/30/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

JUNE 08, 2008

OPERATOR'S NAME: *COB 6/27/08*  
CARRINGTON, KYLE *MC*  
OPERATOR'S NUMBER: 5336  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :

I  
O.L. #: 0  
DEPT/AGENCY: FBK6  
CASE/REPORT: 0  
TEST TYPE: V *MC*  
ALCO TARGET VALUE: .081 ✓  
ALCO S/N: 78895 ✓

### --- BREATH ANALYSIS ---

*80/08/08*

.081 ADJUSTED FOR 29.54 in		
ALCO TARGET	.079	08:40
BLANK TEST	.000	08:40
INTERNAL STANDARD	VERIFIED	08:41
ALCO TV 29.54 in	.080 ✓	08:41
BLANK TEST	.000	08:42
SUBJECT SAMPLE <i>MC</i>	.000 <i>COB 6/27/08</i>	08:42
BLANK TEST	.000	08:43
ALCO TV 29.54 in	.080	08:43
BLANK TEST	.000	08:44

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130121

JUNE 08, 2008 ✓  
TIME 08:37

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 45c  
BAROMETER: 29.54 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ J ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~