VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

Supervisor/Operator Performing the Verification Procedure:

A

Name Kyle Carrington ID# 5336 Date 8-4-08

Agency Fairbanks Airport Police Phone # 474-2555

Instrument Location Fairbanks, AK

B

Alco S/N 78895 Target Value .081 High Pressure 1100 PSI

Alco Test Values .081 (1st Alco) .080 (2nd Alco)

Signature [signature] CWB 8/12/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 8/13/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 AUG 07 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

AUGUST 04, 2008 CDS 8/12/08

OPERATOR'S NAME:
CARRINGTON,KYLE
OPERATOR'S NUMBER: 5336
SUBJECT'S LAST NAME: KC
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
I
O.L. #: 0
DEPT/AGENCY: FBK6
CASE/REPORT: 0 KC
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: 78895

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.68 in

ALCO TARGET	.080	11:52
BLANK TEST	.000	11:53
INTERNAL STANDARD	VERIFIED	11:53
ALCO TV 29.68 in KC	.081	11:54
BLANK TEST	.000	11:55
SUBJECT SAMPLE	.000 CDS 8/12/08	11:55
BLANK TEST	.000	11:56
ALCO TV 29.68 in	.080	11:56
BLANK TEST	.000	11:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

AUGUST 04, 2008
TIME 11:50

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	44c
BAROMETER:	29.68 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~