

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 07 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130121 ✓

Name John Pernel ID# 6260 Date 10/6/09 ^{COB} 10/12/09

A Agency Fairbanks Airport Police & Fire Phone # 474-2555

Instrument Location 5195 Brumbaugh St. Fairbanks, AK 99709

B Alco S/N 57447 Target Value .079 High Pressure 400 psi

Alco Test Values .082 .081
1st Alco 2nd Alco

Signature John Pernel ✓

(OVER)

^{COB}
10/12/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121 —

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

OCTOBER 06, 2009

COB 10/2/09

OPERATOR'S NAME:
PERNELL/JOHN/T
OPERATOR'S NUMBER: 6260
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: UOC
DEPT/AGENCY: FBK6
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: 57447

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.35 in

ALCO TARGET	.000	07:27
BLANK TEST	.000	07:28
INTERNAL STANDARD	VERIFIED	07:28
ALCO TU 30.35 in	.002	07:29
BLANK TEST	.000	07:29
SUBJECT SAMPLE	.000	07:30
BLANK TEST	.000	07:31
ALCO TU 30.35 in	.001	07:31
BLANK TEST	.000	07:32

COB 10/2/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

OCTOBER 06, 2009
TIME 07:24

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 05/29/00
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 45c
BAROMETER: 30.35 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

step

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~■