

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130121 ✓

Name John Pernel ID# 6260 ✓ Date 12-14-10 ✓

A Agency Fairbanks Airport Police Phone # 474-2555

Instrument Location 5195 Brumbaugh Ave. Fairbanks, AK 99709

B Alco S/N X124670 ✓ Target Value .082 ✓ High Pressure 500 psi

Alco Test Values .080 ✓ .079 ✓
1st Alco 2nd Alco

Signature: John Pernel

(OVER)

AMB
12/23/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of JAN, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

DECEMBER 14, 2010

OPERATOR'S NAME:
PERNELL

OPERATOR'S NUMBER: 6260

SUBJECT'S LAST NAME:
UC

SUBJECT'S FIRST NAME/MI :
UC

O.L. #: UC

DEPT/AGENCY: FBK6

CASE/REPORT: UC

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: X124670

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.98 in
ALCO TARGET .082 14:33
BLANK TEST .000 14:34
INTERNAL STANDARD VERIFIED 14:34
ALCO TV 29.98 in .080 14:34
BLANK TEST .000 14:35
SUBJECT SAMPLE .000 14:36
BLANK TEST .000 14:36
ALCO TV 29.98 in .079 14:37
BLANK TEST .000 14:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

DECEMBER 14, 2010
TIME 14:30

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 44c
BAROMETER: 29.98 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~