

**DELTA**

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 22 2009

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130122

Name TIM TUCKWOOD

ID# 3446 Date 10-9-09

Agency Alaska State Troopers

Phone # 895-4800

Instrument Location Delta Courthouse

B

Alco S/N X301596

Target Value .081

High Pressure 500

Alco Test Value Average .082  
1<sup>st</sup> Alco

.081  
2<sup>nd</sup> Alco

Signature

*Tim Tuckwood*

*CB*  
*10/2/09*

(OVER)

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


*Nita J. Bolz*

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 6<sup>th</sup> day of NOV., 2009.

*Carolyn M. Noland*  
(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

*Handwritten mark*

DataMaster cdm S/N 130122 /

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130122

OCTOBER 09, 2009

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CAL

D.L. #: 13345

DEPT/AGENCY: DJTO

CASE/REPORT: 09765

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: X301596

### — BREATH ANALYSIS —

.001 ADJUSTED FOR 29.13 in		
ALCO TARGET	.078	14:08
BLANK TEST	.000	14:09
INTERNAL STANDARD	VERIFIED	14:09
ALCO TU 29.13 in	.002	14:09
BLANK TEST	.000	14:10
SUBJECT SAMPLE	.000	14:10
BLANK TEST	.000	14:11
ALCO TU 29.13 in	.001	14:11
BLANK TEST	.000	14:12

*Handwritten: COS 10/27/09*

*Handwritten: COS 10/27/09*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130122

OCTOBER 09, 2009

TIME 14:06

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.13 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z 1 2 3