VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 01 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130122

A

Name Tim Tuckwood ID# 3446 Date 11-28-09

Agency Alaska State Troopers Phone # 895-4800

Instrument Location Delta Court

B

Alco S/N X301596 Target Value .081 High Pressure 450 psi

Alco Test Value Average .080 (1st Alco) .079 (2nd Alco)

Signature Tim Tuckwood

CBB 12/1/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do not write in the area below)

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

CAROLYN M. NOLAND NOTARY PUBLIC STATE OF ALASKA

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130122

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

NOVEMBER 28, 2009 CDB 12/7/09

OPERATOR'S NAME:
TUCKWOOD TIM
OPERATOR'S NUMBER: 3446
SUBJECT'S LAST NAME:
VERIF
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 12345
DEPT/AGENCY: DJT0
CASE/REPORT: 09876
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X301596

— BREATH ANALYSIS —

.081 ADJUSTED FOR 28.32 in

ALCO TARGET	.076	12:17
BLANK TEST	.000	12:18
INTERNAL STANDARD	VERIFIED	12:18
ALCO TV 28.32 in	.080	12:18
BLANK TEST	.000	12:19
SUBJECT SAMPLE	.000	12:19
BLANK TEST	.000	12:20
ALCO TV 28.34 in	.079	12:20
BLANK TEST	.000	12:21

CDB 12/7/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

NOVEMBER 28, 2009
TIME 12:15

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	48c
BAROMETER:	28.32 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~