

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 03 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130122 ✓	
Name	Kirsten Hansen	ID#	5440 KZH0 ^{DMJ} 8/10	
A Agency		Alaska State Troopers /DJT0	Date	07/29/2010 ✓
Instrument Location		Delta Jct. Court House		
Alco S/N		75602 ✓	Target Value	.081 ✓
High Pressure		925 psi		
B Alco Test Values		0.079 ^{0.080} _{DMJ 8/5/10}	1 st Alco	0.079 ✓
Signature		Kirsten Hansen		
		COB 8/9/10		

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Aug, 2010.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130122

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JULY 29, 2010

OPERATOR'S NAME: HANSEN, KIRSTEN/Z
OPERATOR'S NUMBER: 5440
SUBJECT'S LAST NAME: UOC
SUBJECT'S FIRST NAME/MI: TEST/I
O.L. #: 123456
DEPT/AGENCY: DJT0
CASE/REPORT: 100729
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: 75602

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 28.99 in
ALCO TARGET .078 13:50
BLANK TEST .000 13:51
INTERNAL STANDARD (VERIFIED) 13:51
ALCO TV 28.99 in .080 13:51
BLANK TEST .000 13:52
SUBJECT SAMPLE .000 13:53
BLANK TEST .000 13:53
ALCO TV 28.99 in .079 13:54
BLANK TEST .000 13:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130122

JULY 29, 2010
TIME 13:44

--- DIAGNOSTIC CHECK ---

COMPUTER#: OKAY
PROGRAM#: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER #: 28.99 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR#: OKAY
FILTERS#: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () ~ |