

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAY 31 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130124

Name NICHOLAS YLTIJEMTE ID# 3429 Date 5/27/11 *BMB*

A Agency METLAKATLA POLICE DEPT. Phone # 907 886-4011

Instrument Location METLAKATLA POLICE DEPT.

B Alco S/N X172988 Target Value .078 High Pressure 1000 ps.

Alco Test Values .078 .078
1st Alco 2nd Alco

Signature Nicholas Ylтијemte

(OVER)

BMB
5/31/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. *2011 NJB*

Kyle Klopfer
(Notary Seal Stamp)
Daniel L. Nicholas KYLE KLOPPER
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 27, 2011

OPERATOR'S NAME: *BMB 5/31/11*
YLINIENI/NICHOLAS
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME:
UER
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 0000
DEPT/AGENCY: MET1
CASE/REPORT: 0000
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: X172988

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.01 in		
ALCO TARGET	.078	10:21
BLANK TEST	.000	10:22
INTERNAL STANDARD	VERIFIED	10:22
ALCO TV 30.01 in	.078	10:23
BLANK TEST	.000	10:24
SUBJECT SAMPLE	.000	10:24
BLANK TEST	.000	10:25
ALCO TV 30.03 in	.078	10:25
BLANK TEST	.000	10:26

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 27, 2011
TIME 10:19

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	40c
BREATH TUBE:	42c
BAROMETER:	30.01 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \ J ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * +