

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

AUG 31 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130124

Name NICHOLAS KUENTEMEL ID# 3829 Date 8/29/11

A Agency METLAKATLA POLICE DEPT. Phone (907)886-4011

Instrument Location METLAKATLA POLICE DEPT.

B Alco S/N X172988 Target Value .078 High Pressure 800 psi

Alco Test Values .075 .079
1st Alco 2nd Alco

Signature Nicholas Kuentemel

(OVER)

BMB
9/2/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2011.

Nikki Roth

Nikki Roth, Notary Public
My Commission Expires With Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

AUGUST 29, 2011

OPERATOR'S NAME:
VLINTET/NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME:
UER
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 10000
DEPT/AGENCY: MET1
CASE/REPORT: 00000
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: X172988

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.12 in		
ALCO TARGET	.078	09:58
BLANK TEST	.000	09:59
INTERNAL STANDARD	VERIFIED	09:59
ALCO TV 30.12 in	.075	09:59
BLANK TEST	.000	10:00
SUBJECT SAMPLE	.000	10:01
BLANK TEST	.000	10:01
ALCO TV 30.14 in	.079	10:02
BLANK TEST	.000	10:03

BMB
9/2/11

BMB
9/2/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

AUGUST 29, 2011
TIME 09:56

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
BAROMETER: 30.12 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmnop
qrstuvwxyz{|}~||