VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130124

A

Name NICHOLAS YLINIEMI ID# 3429 Date 2/6/08

Agency METLAKATLA POLICE DEPARTMENT Phone # 886-4011

Instrument Location METLAKATLA POLICE DEPARTMENT

B

Alco S/N X124662 ✓ Target Value .082 ✓ High Pressure 150psi

Alco Test Values .085 (1st Alco) .094 (2nd Alco)

Signature [signature] CDB 2/19/08

(OVER)

(Do Not Write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Nicholas Yliniemi, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORIES

[signature] 02/22/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

FEBRUARY 06, 2008

OPERATOR'S NAME: COB 2/19/08
VLINIENI/NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME:
VER
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 00000
DEPT/AGENCY: METI
CASE/REPORT: 0000
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X124652

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.73 in

ALCO TARGET	.081	09:36
BLANK TEST	.000	09:37
INTERNAL STANDARD	VERIFIED	09:37
ALCO TV 29.73 in	.085	09:38
BLANK TEST	.000	09:39
SUBJECT SAMPLE	.000	09:39
BLANK TEST	.000	09:40
ALCO TV 29.73 in	.084	09:40
BLANK TEST	.000	09:41

COB 2/19/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

FEBRUARY 06, 2008
TIME 09:32

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.73 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~