

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130124</u>
Name		<u>NICHOLAS YLINIEMI</u>	ID# <u>3429</u> Date <u>5/1/08</u>
A	Agency	<u>METLAKATLA POLICE DEPT.</u>	Phone # <u>886-4011</u>
Instrument Location		<u>METLAKATLA POLICE DEPARTMENT</u>	
Alco S/N		<u>78853</u>	Target Value <u>.079</u> High Pressure <u>1200 psi</u>
B	Alco Test Values	<u>.082</u> 1 st Alco	<u>.083</u> 2 nd Alco
Signature		<u>Nicholas Yliniemi</u>	<u>COB</u> <u>5/13/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Nicholas Yliniemi, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

5/28/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 01, 2008

OPERATOR'S NAME: **COB 5/13/08**
VLINIENTI/NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME:
VER
SUBJECT'S FIRST NAME/MI :
OF CAL

O.L. #: 111111
DEPT/AGENCY: MET1
CASE/REPORT: 111111
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: 78853

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.12 in
ALCO TARGET .079 00:49
BLANK TEST .000 00:50
INTERNAL STANDARD VERIFIED 00:50
ALCO TV 30.12 in .002 00:51
BLANK TEST .000 00:51
SUBJECT SAMPLE .000 **COB 5/13/08** 00:52
BLANK TEST .000 00:53
ALCO TV 30.14 in .003 00:53
BLANK TEST .000 00:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

MAY 01, 2008
TIME 08:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.12 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLmnopqrstuvwxyz{|}~^_`'abcde fghi jk lano
pqrstuvwxyz{|}~■