

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124 ✓
FEBRUARY 22, 2010 ✓ **BMB 3/2/10**
OPERATOR'S NAME: PLINTINI-NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME: ✓
SUBJECT'S FIRST NAME/MI: ✓
O.L. #: 99999
DEPT AGENCY: MET1
CASE REPORT: 00000
TEST TYPE: U
ALCO TARGET VALUE: .084 ✓
ALCO S/N: W172940 ✓
--- BREATH ANALYSIS ---
ALCO ADJUSTED FOR 30.24 in ✓
ALCO TARGET .084 00:56
BLANK TEST .000 00:57
INTERNAL STANDARD VERIFIED 00:57
ALCO TO 30.24 in .086 00:57
BLANK TEST .000 00:58
SUBJECT SAMPLE .000 00:58
BLANK TEST .000 00:59
ALCO TO 30.24 in .086 01:00
BLANK TEST .000 01:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124 ✓
FEBRUARY 22, 2010 ✓
TIME 00:50
--- DIAGNOSTIC CHECK ---
COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.24 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY
PRINTER TEST
"H#00" (3) + "-" / 0123456789101112 = 123456789
HIKUNHOPGRSTUVWXYZ "J" "abcde fghijklmno
pqrstuvwxyz{|}~@