

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUN 29 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130124</u> ✓	
Name	<u>NICHOLAS YLINTEMI</u> ID# <u>3429</u> Date <u>6/29/10</u> ✓
A Agency	<u>METLAKATLA POLICE DEPT.</u> Phone # <u>886-4011</u>
Instrument Location <u>METLAKATLA POLICE DEPT.</u>	
B Alco S/N <u>X173012</u> ✓ Target Value <u>.086</u> ✓ High Pressure <u>1200 ps.</u>	
Alco Test Values <u>.088</u> ✓ <u>.088</u> ✓	
1 st Alco 2 nd Alco	
Signature	<u>Nicholas Ylinteri</u> CRS 7/8/10
(OVER)	



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of July 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130124

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

JUNE 25, 2010

OPERATOR'S NAME:
VLINTEMI-NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME:

VER
SUBJECT'S FIRST NAME/MI :
OF CAL

O.L. #: 00000
DEPT/AGENCY: MET1
CASE/REPORT: 09909
TEST TYPE: U
ALCO TARGET VALUE: .086
ALCO S/N: X173012

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 30.20 in
ALCO TARGET .086 13:19
BLANK TEST .000 13:20
INTERNAL STANDARD VERIFIED 13:20
ALCO TV 30.20 in .088 13:20
BLANK TEST .000 13:21
SUBJECT SAMPLE .000 13:21
BLANK TEST .000 13:22
ALCO TV 30.20 in .088 13:22
BLANK TEST .000 13:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

JUNE 25, 2010
TIME 13:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 30.20 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`'abcde fghijklmno
pqrstuvmxyz{|}~■