

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130125</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>GARY J. TELLEP</u>	ID# <u>2281</u> Date <u>4-14-11</u>
A Agency <u>AST D Detachment</u>	Phone # <u>451-5100</u>
Instrument Location <u>1979 Peger Rd. FLKs.</u>	
Alco S/N <u>555124</u> Target Value <u>-080</u> High Pressure <u>400</u>	
B Alco Test Values	<u>-078</u> <u>-079</u>
	1st Alco                  2nd Alco
Signature <u>Gary J. Tellep</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2011. NJB

Kyle Klopfer (Notary Seal Stamp)  
Daniel L. Michotas KYLE KLOPPER  
Notary Public, State of Alaska  
Commission Expires with Office



CRF  
4/21/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

APRIL 14, 2011 CTF 4/21/11

OPERATOR'S NAME:  
TELLEP GARY J  
OPERATOR'S NUMBER: 2281  
SUBJECT'S LAST NAME:  
VERIFICATIONDFCAL  
SUBJECT'S FIRST NAME/MI :  
N  
O.L. #: N  
DEPT/AGENCY: FBK0  
CASE/REPORT: N  
TEST TYPE: U  
ALCO TARGET VALUE: .000  
ALCO S/N: 555124

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.88 in		
ALCO TARGET	.079	03:18
BLANK TEST	.000	03:19
INTERNAL STANDARD	VERIFIED	03:19
ALCO TV 29.88 in	.078	03:20
BLANK TEST	.000	03:20
SUBJECT SAMPLE	.000	03:21
BLANK TEST	.000	03:22
ALCO TV 29.88 in	.079	03:22
BLANK TEST	.000	03:23

CTF 4/21/11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

APRIL 14, 2011  
TIME 03:24

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.88 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢