VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

Supervisor/Operator Performing the Verification Procedure:

A

Name GARY J. Tellep ID# 2281 Date 3-10-08

Agency AST- FbKs Phone # 451 5100

Instrument Location 1979 Peger Rd. FbKs 99701

B

Alco S/N X301600 Target Value .078 High Pressure 1150 psi

Alco Test Value Average .072 (1st Alco) .073 (2nd Alco)

Signature [signature] CAB 3/18/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary Tellep , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 3/18/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORIES

MAR 17 2008

BT9 01/00

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MARCH 10, 2008 CJB 3/18/08

OPERATOR'S NAME:
TELLEP GARY J
OPERATOR'S NUMBER: 2281
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: FBK0
CASE/REPORT: L
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: X301600

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.44 in

ALCO TARGET	.076	20:17
BLANK TEST	.000	20:18
INTERNAL STANDARD	VERIFIED	20:18
ALCO TV 29.44 in	.072	20:18
BLANK TEST	.000	20:19
SUBJECT SAMPLE	.000 CJB 3/18/08	20:20
BLANK TEST	.000	20:20
ALCO TV 29.44 in	.073	20:21
BLANK TEST	.000	20:22

20:18

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MARCH 10, 2008
TIME 20:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.44 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~