

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **Datamaster cdm S/N** 130125 ✓

Name Michael A. Wery ID# 3590 Date 5-8-08 ✓

**A** Agency AST Phone # 451-5100

Instrument Location Fairbanks

**B** Alco S/N X 301600 ✓ Target Value .078 ✓ High Pressure 750

Alco Test Value Average .073 ✓ .074 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Handwritten Signature] COB  
5/14/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael A Wery, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Handwritten Signature] 5/28/08  
Jeanne Swartz Date  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

MAY 08, 2008

OPERATOR'S NAME:

WERY/MICHAEL/A

OPERATOR'S NUMBER: 3590

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CALIBRATION

O.L. #: 0

DEPT/AGENCY: FBK0

CASE/REPORT:

TEST TYPE: T

ALCO TARGET VALUE: .078

ALCO S/N: X301600

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.77 in  
ALCO TARGET .077 08:18  
BLANK TEST .000 08:19  
INTERNAL STANDARD VERIFIED 08:19  
ALCO TV 29.77 in .073 08:19  
BLANK TEST .000 08:20  
SUBJECT SAMPLE .000 08:21  
BLANK TEST .000 08:22  
ALCO TV 29.77 in .074 08:22  
BLANK TEST .000 08:23

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

MAY 08, 2008

TIME 08:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.77 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()\*+,-./0123456789:;<=?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghi.jklmno  
pqrstuvwxy{|}~■