

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

T

T

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 100012

JULY 05, 2003

OPERATOR'S NAME:
GOEDEN, CHAD J
OPERATOR'S NUMBER: 3513
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI:
OF DAL

O.L.: 41 XXXX
DEPT./AGENCY: FB/0
CASE/REPORT: XXXX
TEST TARGET: 0
ALCO TARGET VALUE: 0.00
ALCO S/N: W001531

--- BREATH ANALYSIS ---

| | | |
|-----------------------------|----------|-------|
| TEST INDICATED FOR 09-00 in | | |
| ALCO TEST | 0.02 | 18:11 |
| BLANK TEST | 0.00 | 18:13 |
| INTERNAL STANDARD | VERIFIED | 18:13 |
| ALCO TO 29.90 in | 0.76 | 18:13 |
| BLANK TEST | 0.00 | 18:14 |
| SUBJECT SAMPLE | 0.00 | 18:15 |
| BLANK TEST | 0.00 | 18:15 |
| ALCO TO 29.90 in | 0.77 | 18:16 |
| BLANK TEST | 0.00 | 18:16 |

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 100012

JULY 05, 2003
TIME 18:19

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 15c
SPEECH TUBE: 11c
BAROMETER: 29.90 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO TO 29.90 in 0.76
ALCO TO 29.90 in 0.77



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 22 2009

| | | | |
|--|------------------|--|---|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | <u>130125</u> ✓ |
| Name | <u>C. Goeden</u> | ID# | <u>3513</u> Date <u>7-5-09</u> |
| A | Agency | <u>AST</u> | Phone # <u>451-5100</u> |
| Instrument Location | | <u>Fairbanks</u> ✓ | |
| B | Alco S/N | <u>X301591</u> ✓ | Target Value <u>0.083</u> ^{0.083} 0.082 High Pressure <u>1250 PSI</u> |
| Alco Test Values | | <u>0.076</u> ✓ 1 st Alco | <u>0.077</u> ✓ 2 nd Alco |
| Signature | | <u>[Signature]</u> ✓ | |

088
7/23/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of July, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03

