VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

A

Supervisor/Operator Performing the Verification Procedure:

Name Michael A. Wery ID# 3590 Date 7-14-08

7-24-08 CWW

Agency AST Phone # 451-5100

Instrument Location Fairbanks

B

Alco S/N X172971 Target Value .079 High Pressure 1050

Alco Test Value Average .074 (1st Alco) .076 (2nd Alco)

Signature [signature] 7/25/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael A Wery, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 7/28/08

Jeanne Swartz — Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 01/00

JUL 21 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

JULY 14, 2008 COB 7/25/08

OPERATOR'S NAME:
WERY/MICHAEL/A
OPERATOR'S NUMBER: 3590
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CALIBRATION
O.L. #: 0
DEPT/AGENCY: FBK0
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X172971

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.85 in

ALCO TARGET	.078	20:46
BLANK TEST	.000	20:47
INTERNAL STANDARD	VERIFIED	20:47
ALCO TV 29.85 in	.074	20:47
BLANK TEST	.000	20:48
SUBJECT SAMPLE	.000	20:49
BLANK TEST	.000	20:50
ALCO TV 29.83 in	.076	20:50
BLANK TEST	.000	20:51

COB 7/25/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

JULY 14, 2008
TIME 20:43

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.85 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~