

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

NOV 13 2009

NOV 13 2009 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:
Datamaster cdm S/N 130125

Name GARY J. TELLOP ID# 2281 Date 11-12-09

Agency AST D DETACHMENT Phone # 451-5100

Instrument Location 1979 PEGER RD. FBKS. 99701

Alco S/N X172966 Target Value .079 High Pressure 1125

Alco Test Value Average .080 .083
1st Alco 2nd Alco

Signature Gary J. Tellop

(OVER)

BMB
12/31/09

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows.
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of January 2010
2009

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130125

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

NOVEMBER 12, 2009

OPERATOR'S NAME:

TELLEP GARY J

OPERATOR'S NUMBER: 2261

SUBJECT'S LAST NAME:

VERIFICATION/FOCAL

SUBJECT'S FIRST NAME/MI:

N

O.L. #: N

DEPT/AGENCY: FB/KO

CASE/REPORT: N

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: W172966

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 28.92 in ✓
ALCO TARGET .076 03:29
BLANK TEST .000 03:30
INTERNAL STANDARD VERIFIED 03:30
ALCO TO 28.92 in .000 03:31
BLANK TEST .000 03:31
SUBJECT SAMPLE .000 03:33
BLANK TEST .000 03:34
ALCO TO 28.92 in .003 03:34
BLANK TEST .000 03:35

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

NOVEMBER 12, 2009

TIME 03:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 28.92 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;K=>@ABCDEF
HIJKL/MNOPQRSTUVWXYZ\]^_`abcde fghijklano
pqrstuvwxyz{|}~||