

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 19 2010

Datamaster cdm S/N 130125

Supervisor/Operator Performing the Verification Procedure:

Name Anthony M. Wiles ID# 6877 Date 1/9/10

A Agency AST-D Detachment Phone # 451-5100

Instrument Location 1979 PEGER Rd FBRs, 99701

---

Alco S/N X172966 Target Value .079 High Pressure 400 psi

B Alco Test Value Average .083 .084

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Anthony Wiles CWJ  
1/22/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of January 2010.

Carolyn M. Nofand (Notary Seal Stamp)  
Carolyn M. Nofand  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130125

(CONTINUED FROM FRONT PAGE)

C

**TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

JANUARY 09, 2010

OPERATOR'S NAME: WILES ANTHONY M  
OPERATOR'S NUMBER: 6877  
SUBJECT'S LAST NAME: VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI: N

O.L. #: N  
DEPT/AGENCY: FBK0  
CASE/REPORT: N  
TEST TYPE: V  
ALCO TARGET VALUE: .079  
ALCO S/N: 1172966

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.70 in  
ALCO TARGET .079 06:16  
BLANK TEST .000 06:17  
INTERNAL STANDARD VERIFIED 06:17  
ALCO TU 29.70 in .083 06:17  
BLANK TEST .000 06:18  
SUBJECT SAMPLE .000 06:19  
BLANK TEST .000 06:20  
ALCO TU 29.70 in .083 06:20  
BLANK TEST .000 06:21

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

JANUARY 09, 2010  
TIME 06:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 2/1/2001  
HEATERS  
SAMPLE CHAMBER: 45c  
BREATH TUBE: 47c  
BAROMETER: 29.70 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINT TEST

ALCO TU 29.70 in .083 06:20  
ALCO TU 29.70 in .083 06:20  
ALCO TU 29.70 in .083 06:20