VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 10 2010

Datamaster cdm S/N 130125

Supervisor/Operator Performing the Verification Procedure:

A

Name Anthony M. Wiles **ID#** 6877 **Date** 5-15-10

Agency AST **Phone #** 451-5100

Instrument Location 1979 Peger Rd.

B

Alco S/N 105859 **Target Value** 0.082 **High Pressure** 925 psi

Alco Test Value Average .076 (1st Alco) .077 (2nd Alco)

Signature Anthony Wiles

5/24/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of June, 2010.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 01/00

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C **TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MAY 15, 2010

OPERATOR'S NAME:
WILES ANTHONY M
OPERATOR'S NUMBER: 6877
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
N
O.L. #: N
DEPT/AGENCY: FBKO
CASE/REPORT: N
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: 65859

BMB 5/24/10

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.41 in

ALCO TARGET	.080	08:38
BLANK TEST	.000	08:39
INTERNAL STANDARD	VERIFIED	08:39
ALCO TV 29.41 in	.076	08:39
BLANK TEST	.000	08:40
SUBJECT SAMPLE	.000	08:41
BLANK TEST	.000	08:41
ALCO TV 29.41 in	.077	08:42
BLANK TEST	.000	08:42

BMB 5/24/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MAY 15, 2010
TIME 08:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	52c
BREATH TUBE:	42c
BAROMETER:	29.41 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~