

AUG 26 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N 130125** ✓

Name GARY J. Tellep ID# 2281 ✓ Date 8/18/2010 <sup>8/30/10</sup>

Agency AST Ft D Det. Phone # 451.5100

Instrument Location 1979 Peyer Rd., Ft 99701

---

Alco S/N X 172933 ✓ Target Value - 079 ✓ High Pressure 1300

Alco Test Values -074 1<sup>st</sup> Alco -075 2<sup>nd</sup> Alco

Signature Gary J. Tellep

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 16th day of Sept, 2010.

Carolyn M. Noland  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

AUGUST 18, 2010

OPERATOR'S NAME:  
SUPERVISOR

OPERATOR'S NUMBER: 2281

SUBJECT'S LAST NAME:

VERIFICATION/PCAL

SUBJECT'S FIRST NAME/MI:

N

O.L. #: N

DEPT/AGENCY: FBK0

CASE/REPORT: N

TEST TYPE: U

BLOOD TARGET VALUE: .075

BLOOD S/N: X172933

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.55 in  
BLOOD TARGET .075 15:23  
BLANK TEST .000 15:24  
INTERNAL STANDARD ESTIMATE 15:24  
BLOOD TO 29.55 in .074 15:25  
BLANK TEST 15:26  
SUBJECT SAMPLE 15:26  
BLANK TEST 15:27  
BLOOD TO 29.55 in .075 15:27  
BLANK TEST 15:28

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

AUGUST 18, 2010  
TIME 06:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 42c

BAROMETER: 29.55 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKL@PQRSTU@VWXYZ{|}~' abcdefghijklmnop  
qrsstuvwxyz{|}~||