

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130125 ✓

Name Anthony M. Wiles ID# 6877 Date 10-28-2010 ✓

A Agency AST Phone # 451-5100

Instrument Location 1979 Peger Rd. FBLS, AK 99701

---

B Alco S/N 555124 ✓ Target Value 0.080 ✓ High Pressure 1200 psi

Alco Test Values .0749 .0759  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Anthony Wiles

(OVER)

BMB  
12/7/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 7<sup>th</sup> day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)  
Carolyn M. Nolan  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

OCTOBER 28, 2010

OPERATOR'S NAME: *BMB 12/7/10*  
WILES/ANTHONY/MICHAEL  
OPERATOR'S NUMBER: 6877  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
N  
D.L. #: N  
DEPT/AGENCY: FD/3  
CASE/REPORT: N  
TEST TYPE: U  
SLED TARGET VALUE: .000  
ALCO S/N: 555124

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 28.97 in		
ALCO TARGET	.077	15:58
BLANK TEST	.000	15:59
INTERNAL STANDARD	VERIFIED	15:59
ALCO TV 28.97 in	.074	16:00
BLANK TEST	.000	16:00
SUBJECT SAMPLE	.000	16:01
BLANK TEST	.000	16:01
ALCO TV 28.95 in	.075	16:02
BLANK TEST	.000	16:03

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125

OCTOBER 28, 2010  
TIME 16:03

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 43c  
BREATH TUBE: 41c  
BAROMETER: 28.95 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`ab`cde+ghi,jk lano  
pqrstuvwxyz{|}~#