

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130126 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name Michael Farmer ID# 6450 ✓ Date 3/5/11 ✓

A Agency Anchorage Airport Police Phone # 907-266-2411
 Instrument Location Anch. Airport Police DUI Processing Room

B Alco S/N X301583 ✓ Target Value .078 ✓ High Pressure 1250psi

Alco Test Values .076 ✓ 1st Alco .076 ✓ 2nd Alco

Signature [Signature] 54

(OVER)

[Handwritten initials]
4/11/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 1st day of APRIL, 2011

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130126 *SMB 4/1/11*

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

[]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 05, 2011 *SMB 4/1/11*

OPERATOR'S NAME: FARMER, MICHAELA
OPERATOR'S NUMBER: 6458
SUBJECT'S LAST NAME: UOC
SUBJECT'S FIRST NAME/MI: UOC
O.L. #: 1234567
DEPT/AGENCY: ANC6
CASE/REPORT: 11-1234
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: X301503

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.07 in

ALCO TARGET	.078	19:10
BLANK TEST	.000	19:11
INTERNAL STANDARD	VERIFIED	19:11
ALCO TU 30.07 in	.076	19:12
BLANK TEST	.000	19:13
SUBJECT SAMPLE	.000	19:13
BLANK TEST	.000	19:14
ALCO TU 30.07 in	.076	19:14
BLANK TEST	.000	19:15

SMB 4/1/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 05, 2011
TIME 19:16

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.07 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()+-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxy{|}~