

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130126</u> ✓
Name: <u>mike Farmer</u>	ID#: <u>6450</u>	Date: <u>1/16/09</u> ✓
A Agency: <u>Anchorage Air Port Police</u>		Phone #: <u>266-2411</u>
Instrument Location: <u>Anch Air Port Police DUI Process Room</u> ✓		
Alco S/N: <u>X172968</u> ✓		Target Value: <u>.081</u> ✓
		High Pressure: <u>800 300</u>
B Alco Test Value Average:		
<u>.081</u> ✓ 1 st Alco		<u>.081</u> ✓ 2 nd Alco
Signature: <u>[Signature]</u> ✓		<u>CRB</u> <u>1/23/09</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that,


(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael A Farmer, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

2/3/09
Date

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

JANUARY 16, 2009 ✓

OPERATOR'S NAME:

FARNER

OPERATOR'S NUMBER: 6450

SUBJECT'S LAST NAME:

FARNER

SUBJECT'S FIRST NAME/MI :

MIKE

O.L. #: 123456

DEPT/AGENCY: ANCS

CASE/REPORT: 09-01

TEST TYPE: U

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X172968 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.78 in
ALCO TARGET .080 06:50
BLANK TEST .000 06:51
INTERNAL STANDARD VERIFIED 06:51
ALCO TV 29.78 in .081 ✓ 06:52
BLANK TEST .000 06:52
SUBJECT SAMPLE .080 06:53
BLANK TEST .000 06:54
ALCO TV 29.78 in .081 ✓ 06:54
BLANK TEST .000 06:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

JANUARY 16, 2009 ✓

TIME 06:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefghijklmnop
qrsuvwxyz{|}~