

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

MAR 09 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130126</u> ✓
Name: <u>Derrick Westlake</u> ID#: <u>5574</u> Date: <u>2/24/09</u> ✓		
A	Agency: <u>Anchorage Airport Police</u> Phone #: <u>266-2411</u>	
Instrument Location: <u>6040 Dehavilland Ave., Anchorage 99502</u> ✓		
Alco S/N: <u>X124656</u> ✓ Target Value: <u>.079</u> ✓ High Pressure: <u>1350 psi</u>		
B	Alco Test Value Average: <u>.079</u> ✓ 1 st Alco	<u>.078</u> ✓ 2 nd Alco
Signature: <u>Derrick Westlake</u> ✓		<u>CSB</u> <u>3/24/09</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that:

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derrick M Westlake, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

3/24/09

Date

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

FEBRUARY 24, 2009

OPERATOR'S NAME:
WESTLAKE/DERRICK/M
OPERATOR'S NUMBER: 5574
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 123456
DEPT/AGENCY: ANC6
CASE/REPORT: TEST
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X124656

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.24 in		
ALCO TARGET	.079	11:59
BLANK TEST	.000	12:01
INTERNAL STANDARD	VERIFIED	12:01
ALCO TV 30.24 in	.079	12:01
BLANK TEST	.000	12:02
SUBJECT SAMPLE	.000	12:03
BLANK TEST	.000	12:04
ALCO TV 30.24 in	.079	12:04
BLANK TEST	.000	12:05

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

FEBRUARY 24, 2009
TIME 12:00

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.24 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLNPQRSTUWXYZ[\]^_`abcdefgijklmno
pqrsuvwxyz{|}~