

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126		
Supervisor/Operator Performing the Verification Procedure:		
Name: Lt. Eric Saffell	ID#: 2853	Date: 03/27/08 ✓ A
Agency: Anchorage Airport Police	Phone #: 266-2579	
Instrument Location: Anchorage Airport Police Department		
Alco S/N: X172968	Target Value: .081	High Pressure: 1400 psi
B		
Alco Test Value Average:	.082 ✓ 1 st Alco	.082 ✓ 2 nd Alco
Signature: <u>LT Eric Saffell</u>	COB 4/8/08	
(OVER)		

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Eric Saffell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/12/08

Date

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 27, 2008 *COB 4/8/08*

OPERATOR'S NAME: SAFFELL/ERIC
OPERATOR'S NUMBER: 2853
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI: L
O.L. #: L
DEPT/AGENCY: ANCG
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: X172968

--- BREATH ANALYSIS ---

101

.081 ADJUSTED FOR 30.16 in		
ALCO TARGET	.081	17:32
BLANK TEST	.000	17:33
INTERNAL STANDARD	VERIFIED	17:33
ALCO TV 30.14 in	.082	17:34
BLANK TEST	.000	17:35
SUBJECT SAMPLE	.080 <i>COB 4/8/08</i>	17:35
BLANK TEST	.000 <i>4/8/08</i>	17:36
ALCO TV 30.14 in	.082	17:37
BLANK TEST	.000	17:38

30

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 27, 2008
TIME 17:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.16 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "# \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~