

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

MAY 21 2009

DataMaster cdm S/N: 130126		
Supervisor/Operator Performing the Verification Procedure:		
Name: Lt. Eric Saffell	ID#: 2853	Date: 04/22/09
A		
Agency: Anchorage Airport Police	Phone #: 266-2579	
Instrument Location: Anchorage Airport Police Department		
B		
Alco S/N: X124656	Target Value: .079	High Pressure: 1000 psi
Alco Test Value Average:	.080	.080
	1 st Alco	2 nd Alco
Signature: <u>Lt Eric Saffell</u>		
(OVER)		

BMB
5/22/09


I, Nita J. Bolz, after being first duly sworn, depose (Do Not write in this area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly reviewed activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27th day of May 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

DIAGNOSTIC CHECK

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

APRIL 22, 2009

BMB
5/22/09

OPERATOR'S NAME:
SAFFELL/ERIC
OPERATOR'S NUMBER: 2853
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: ANC6
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X124656 ✓

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

APRIL 22, 2009 ✓
TIME 20:13

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.46 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

.079 ADJUSTED FOR 30.46 in
ALCO TARGET .000 20:17
BLANK TEST .000 20:18
INTERNAL STANDARD VERIFIED 20:18
ALCO TV 30.46 in .000 ✓ 20:18
BLANK TEST .000 20:19
SUBJECT SAMPLE BMB 20:20
BLANK TEST 5/22/09 .000 20:20
ALCO TV 30.46 in .000 ✓ 20:21
BLANK TEST .000 20:22

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~