

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130126

MAY 24, 2008 *CO2 3/08*

OPERATOR'S NAME:
FARMEHIKEA
OPERATOR'S NUMBER: 6450
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: UOC
DEPT/AGENCY: ANC6
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: K172968

--- BREATH ANALYSIS ---

40/21

.081 ADJUSTED FOR 30.29 in		
ALCO TARGET	.082	07:53
BLANK TEST	.000	07:54
INTERNAL STANDARD	VERIFIED	07:55
ALCO TU 30.29 in	.082	07:55
BLANK TEST	.000	07:56
SUBJECT SAMPLE	.000 <i>CO2 3/08</i>	07:56
BLANK TEST	.000	07:57
ALCO TU 30.29 in	.082	07:57
BLANK TEST	.000	07:58

40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130126

MAY 24, 2008
TIME 08:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "\$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrsuvwxyz{|}~#