

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

NOVEMBER 16, 2008 ✓ *COB 11/20/08*

OPERATOR'S NAME:

FARMER

OPERATOR'S NUMBER: 6458

SUBJECT'S LAST NAME:

FARMER

SUBJECT'S FIRST NAME/MI :

MIKE

D.L. #: TEST

DEPT/AGENCY: ANC6

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X172968 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.22 in		
ALCO TARGET	.081	09:27
BLANK TEST	.000	09:28
INTERNAL STANDARD	VERIFIED	09:28
ALCO TV 30.24 in	.082 ✓	09:29
BLANK TEST	.000	09:30
SUBJECT SAMPLE <i>COB 11/20/08</i>	.000	09:30
BLANK TEST	.000	09:31
ALCO TV 30.24 in	.082 ✓	09:31
BLANK TEST	.000	09:32

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

NOVEMBER 16, 2008 ✓
TIME 09:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.22 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { } ~ ■