

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

APR 14 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130126</u>
Name: <u>Michael Farmer</u>		ID#: <u>6450</u> Date: <u>4/3/10</u>
A	Agency: <u>Anchorage Airport Police</u>	Phone #: <u>266-2411</u>
Instrument Location: <u>A.A.P. DUI Processing Room</u>		
Alco S/N: <u>X172961</u>		Target Value: <u>.079</u> High Pressure: <u>1000psi</u>
B	Alco Test Value Average:	<u>.076</u> ✓ <u>.076</u> ✓ 1 st Alco 2 nd Alco
Signature: <u>[Signature]</u> SA MAG		
(OVER)		

BMB
4/15/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of April, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT6 9/95



VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

APRIL 03, 2010 ✓

OPERATOR'S NAME: *BMB 4/15/10*
FARMER, MICHAEL, A
OPERATOR'S NUMBER: 6450
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567
DEPT/AGENCY: ANCS
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X172961 ✓

BREATH ANALYSIS

BMB 4/15/10
.079 ADJUSTED FOR 29.29 in
ALCO TARGET .077 11:13
BLANK TEST .000 11:14
INTERNAL STANDARD VERIFIED 11:14
ALCO TV 29.29 in .076 ✓ 11:14
BLANK TEST .000 11:15
SUBJECT SAMPLE .000 11:16
BLANK TEST .000 11:17
ALCO TV 29.29 in .076 ✓ 11:17
BLANK TEST .000 11:18

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

APRIL 03, 2010 ✓
TIME 11:10 ✓

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL MNOPQRSTUVWXYZ[\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~■