

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

MAY 12 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130127</u>	
Supervisor/Operator Performing the Verification Procedure: Name <u>Christopher Umbs</u> ID# <u>3669</u> Date <u>5/6/11</u>	
A	Agency <u>Alaska State Troopers</u> Phone # <u>465-4004</u>
	Instrument Location <u>Anchorage</u>
Alco S/N <u>70034</u> Target Value <u>.079</u> High Pressure <u>150 psi</u>	
B	Alco Test Values <u>.082</u> <u>.082</u> 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco
	Signature <u>[Handwritten Signature]</u>

(OVER)

*UMB*  
*5/13/11*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. *NJB*

[Handwritten Signature] (Notary Seal Stamp)  
Daniel L. Nicholas KYLE KLOPPER  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130127 *EMB 5/13/11*

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130127

MAY 06, 2011

OPERATOR'S NAME: *EMB 5/13/11*  
UMBS/CHRISTOPHER/J  
OPERATOR'S NUMBER: 3669  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
VOC  
O.L. #: VOC  
DEPT/AGENCY: ANSO  
CASE/REPORT: VOC  
TEST TYPE: U  
ALCO TARGET VALUE: .079  
ALCO S/N: 70034

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.90 in		
ALCO TARGET	.070	09:09
BLANK TEST	.000	09:10
INTERNAL STANDARD	VERIFIED	09:10
ALCO TV 29.90 in	.082	09:10
BLANK TEST	.000	09:11
SUBJECT SAMPLE	.000	09:11
BLANK TEST	.000	09:12
ALCO TV 29.90 in	.082	09:13
BLANK TEST	.000	09:13



## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130127

MAY 06, 2011  
TIME 09:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09

HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 42c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=?@ABCDEFGHIJKLMN  
OPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~