

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

NOV 07 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130127</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>Christopher Umbs</u>	ID# <u>3069</u> Date <u>11/3/11</u>	
A Agency <u>Alaska State Troopers</u>	Phone # <u>465-4004</u>	
Instrument Location <u>Anchorage</u>		
B Alco S/N <u>70034</u> Target Value <u>.079</u> High Pressure <u>90psi</u>		
Alco Test Values	<u>.083</u> 1st Alco	<u>.083</u> 2 <sup>nd</sup> Alco
Signature		
COB 11/2/11		

(OVER)


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 2011 AK

 (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

100 COB "11/7/11"  
130127  
COB "11/7/11"

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130127

NOVEMBER 03, 2011

OPERATOR'S NAME:

UNBS/CHRISTOPHER/J

OPERATOR'S NUMBER: 3669

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: AST

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: 70034

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.20 in  
ALCO TARGET .079 08:21  
BLANK TEST .000 08:22  
INTERNAL STANDARD VERIFIED 08:23  
ALCO TO 30.20 in .083 08:23  
BLANK TEST .000 08:24  
SUBJECT SAMPLE .000 08:24  
BLANK TEST .000 08:25  
ALCO TO 30.20 in .083 08:25  
BLANK TEST .000 08:26

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130127

NOVEMBER 03, 2011

TIME 08:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.20 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~