


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAY 25 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130127 ✓			
Name	Christopher Umbs	ID#	3669			
		Date	05/22/2010			
A	Agency	Alaska State Troopers	Phone # (907) 465-4004			
	Instrument Location	Angoon Public Safety Office				
Alco S/N		70034 ✓	Target Value	.079 ✓	High Pressure	400 psi
Alco Test Values		.081 ✓		.081 ✓		
		1 st Alco		2 nd Alco		
Signature					CORB 6/1/10	

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of June, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130127 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127 ✓

MAY 22, 2010 ✓

OPERATOR'S NAME: *COB 6/1/10*
UMES/CHRISTOPHER/J
OPERATOR'S NUMBER: 3669
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: ANGO
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 70034 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.11 in ✓
ALCO TARGET .079 09:20
BLANK TEST .000 09:21
INTERNAL STANDARD VERIFIED 09:21
ALCO TV 30.12 in .081 ✓ 09:21
BLANK TEST .000 09:22
SUBJECT SAMPLE *COB 6/1/10* .000 09:22
BLANK TEST .000 09:23
ALCO TV 30.11 in .081 ✓ 09:23
BLANK TEST .000 09:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127 ✓

MAY 22, 2010 ✓
TIME 09:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 30.11 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~