


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program DEC 29 2010


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130127</u>
Name		<u>CHRISTOPHER UMBS</u>	ID# <u>3669</u> Date <u>12/23/10</u>
A	Agency	<u>ALASKA STATE TROOPERS</u>	Phone # <u>465-4000</u>
Instrument Location		<u>ANGOON</u>	
Alco S/N		<u>70034</u>	Target Value <u>.079</u> High Pressure <u>300 psi</u>
Alco Test Values		<u>.081</u> 1st Alco	<u>.080</u> 2 nd Alco
Signature			
		CUB 12/30/10	

(OVER)


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of MAR, ~~2010~~ 2011


(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130127
COB 12/30/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127

DECEMBER 23, 2010

OPERATOR'S NAME:
UMBS/CHRISTOPHER/J
OPERATOR'S NUMBER: 3669
SUBJECT'S LAST NAME:

UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC
DEPT/AGENCY: ANSO
CASE/REPORT: UOC
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: 70034

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.80 in
ALCO TARGET .078 10:45
BLANK TEST .000 10:46
INTERNAL STANDARD VERIFIED 10:46
ALCO TV 29.80 in .081 10:46
BLANK TEST .000 10:47
SUBJECT SAMPLE .000 10:48
BLANK TEST .000 10:48
ALCO TV 29.80 in .080 10:49
BLANK TEST .000 10:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127

DECEMBER 23, 2010
TIME 10:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 29.80 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢