

JUN 22 2009

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130128</u> ✓
Name	<u>Brandi Barnett</u>	ID#	<u>7038</u> Date <u>6/17/2009</u> ✓
A	Agency	<u>SCDL</u>	Phone # <u>269-5695</u>
	Instrument Location	<u>SCDL</u>	
B	Alco S/N	<u>SCOTT08</u> ✓	Target Value <u>0.080</u> High Pressure <u>700 psi</u>
	Alco Test Value Average	<u>0.077</u> ✓	<u>0.078</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature	<u>Brandi M. Barnett</u> ✓		<u>COB J 6/23/09</u> <u>6/23/09</u>

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and testify as follows (DataMaster follows area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of June, 2009.

Carolyn M. Noland (Notary Seal Stamp)  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130128 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130128

JUNE 17, 2009 ✓ *CB*

OPERATOR'S NAME:

BARNETT, B

OPERATOR'S NUMBER: 7038

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: AAA1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: 0.000 ✓

ALCO S/N: SCOTT08

— BREATH ANALYSIS —

.000 ADJUSTED FOR 29.72 in		
ALCO TARGET	.079	07:32
BLANK TEST	.000	07:33
INTERNAL STANDARD	VERIFIED	07:33
ALCO TU 29.72 in	.077 ✓	07:33
BLANK TEST	.000	07:34
SUBJECT SAMPLE	.000	07:34
BLANK TEST	.000	07:35
ALCO TU 29.72 in <i>CB</i>	.078 ✓	07:35
BLANK TEST	.000	07:36

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130128

JUNE 17, 2009 ✓  
TIME 07:30

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	29.72 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFB  
HIJKL/MNOPQRSTUWXYZ\]^\_`'abcde fghi jklmno  
pqrstuvwxyz{|}~!@