

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name	<u>JAMES MICHELS</u> ID# <u>6502</u> Date <u>1-14-11</u>
A Agency	<u>USIB POLICE DEPT</u> Phone # <u>852-0311</u>
Instrument Location <u>BARROW - SQUAD ROOM</u>	
DataMaster cdm S/N <u>130129</u> ✓	
B Alco S/N	<u>77675</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>1250 psc</u>
Alco Test Values	
1st Alco	<u>.088</u> ✓
2nd Alco	<u>.086</u> ✓
Signature <u>SGT [Signature] SGT. MICHELS</u>	

(OVER)

BMB
2/8/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of MAR, 2011 AM

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129 ^{BMB} 2/8/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

JANUARY 14, 2011

OPERATOR'S NAME: MICHELS/JAMES/D
OPERATOR'S NUMBER: 6502
SUBJECT'S LAST NAME: VOC
SUBJECT'S FIRST NAME/MI : VOC
O.L. #: VOC
DEPT/AGENCY: DRVI
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .004
ALCO S/N: 77675

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 31.03 in

ALCO TARGET	.007	13:18
BLANK TEST	.000	13:19
INTERNAL STANDARD	VERIFIED	13:19
ALCO TV 31.03 in	.008	13:19
BLANK TEST	.000	13:20
SUBJECT SAMPLE	.000	13:21
BLANK TEST	.000	13:22
ALCO TV 31.03 in	.006	13:22
BLANK TEST	.000	13:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

JANUARY 14, 2011
TIME 13:16

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	31.02 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD =	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLmnopqrstuvwxyz{|}~_`abcde fghi jk lmnop
pqrstuvwxyz{|}~`