VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129

Supervisor/Operator Performing the Verification Procedure:

A

Name SGT JAMES MICHIELS ID# 6502 Date 4-18-11

Agency USBPN Phone # 852-0311

Instrument Location BARROW / SQUAD ROOM

B

Alco S/N 77675 Target Value .084 High Pressure 900 psi

Alco Test Values .086 (1st Alco) .087 (2nd Alco)

Signature 575

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26 day of APRIL, ~~2010~~ 2011 KK.

(Notary Seal Stamp)
~~Daniel L. Nicholas~~ KYLE KLOPFER
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

KYLE KLOPFER NOTARY PUBLIC STATE OF ALASKA

CRF 4/21/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129 CRF 4/21/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

APRIL 18, 2011 CRF 4/21/11

OPERATOR'S NAME:
MICHELS/JAMES/D
OPERATOR'S NUMBER: 6502
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: DR01
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .084
ALCO S/N: 77675

— BREATH ANALYSIS —

.084 ADJUSTED FOR 30.45 in

ALCO TARGET	.085	04:24
BLANK TEST	.000	04:25
INTERNAL STANDARD	VERIFIED	04:25
ALCO TV 30.45 in	.086	04:26
BLANK TEST	.000	04:26
SUBJECT SAMPLE	.000	04:27
BLANK TEST	.000	04:28
ALCO TV 30.45 in	.087	04:28
BLANK TEST	.000	04:29

CRF 4/21/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

APRIL 18, 2011
TIME 04:22

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.45 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~