

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 16 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130129</u> ✓
Name	<u>William Rapson</u>	ID# <u>6208</u> Date <u>3/03/09</u> ✓
A Agency	<u>WAS. LLA Police Department</u>	Phone # <u>352-5401</u>
Instrument Location	<u>WAS. LLA P.D.</u> ✓	
B Alco S/N	<u>X129625</u>	Target Value <u>.081</u> ✓ High Pressure <u>1275 PSI</u>
Alco Test Value Average	<u>.080</u> ✓	<u>.081</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature	<u>Will R.</u>	<u>CRB</u> <u>7/16/09</u>

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of July, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129

Handwritten: 130129  
7/10/09

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY

### Breath Test Record

Serial Number 130129 Model DataMaster

Location WAS1 Wasilla Police Dept.

Handwritten: CBP 7/10/09

### Subject Information

Last Name Voc O.L. # 1234567

First Name Voc

### Operator Information

Operator Name Rapson/William/C Case VOC

Operator # 6208 Department WAS1

### Test Results

Alco Target .081 Alco Serial 124625

Nominal 0.079 Barometer 29.41 Test Type V

Dry Gas 1 0.080 Time: 17:26:38 Test Date 03/23/2009

Subject Alcohol 0.000 Time: 17:27:55 Test Time 17:25:24

Dry Gas 2 0.081 Time: 17:28:54 Report Y

## ALASKA DEPARTMENT OF PUBLIC SAFETY

### Diagnostic Test Record

Serial Number 130129 Model DataMaster

Location WAS1 Wasilla Police Dept.

### Diagnostic Test Results

Test Date 03/23/2009 Computer Okay

Test Time 17:22:42 Program Okay

Software Date 01/08/2009

Sample Chamber (C) 48

Breath Tube (C) 41

Barometer (in) 29.41

Flow Detector Okay

High Speed Pump Okay

Detector Okay

Filter Okay

Quartz Okay

Calibration Okay