



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129

BMB  
10/1/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130129

SEPTEMBER 28, 2010

OPERATOR'S NAME:

NICHOLS/JAMES/D

OPERATOR'S NUMBER: 6502

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: DRUI

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X301568

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.73 in

ALCO TARGET	.003	15:53
BLANK TEST	.000	15:54
INTERNAL STANDARD	VERIFIED	15:54
ALCO TV 29.73 in	.003	15:55
BLANK TEST	.000	15:55
SUBJECT SAMPLE	.000	15:56
BLANK TEST	.000	15:57
ALCO TV 29.73 in	.001	15:57
BLANK TEST	.000	15:58

BMB  
10/1/10

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130129

SEPTEMBER 28, 2010

TIME 15:59

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.73 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~`