

JAN 13 2011
2011
2010
3/8/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130130 ✓


Name Jeremy Conkling ID# 6715 Date 01.05.11 ✓

A Agency VPD Phone # 835 4560

Instrument Location VPD Jail

B Alco S/N 57324 ✓ Target Value .087 ✓ High Pressure 295

Alco Test Values .086 ✓ 1st Alco .088 ✓ 2nd Alco

Signature  271


(OVER)

8MB
2/8/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of MAR, 2011 ²⁰¹¹ ₂₀₁₀ ^{an}

 (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JANUARY 05, 2011

OPERATOR'S NAME:

CONKLING/JEREMY/W

OPERATOR'S NUMBER: 6715

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: UDZ1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .087

ALCO S/N: 57324

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.55 in
ALCO TARGET .085 01:07
BLANK TEST .000 01:07
INTERNAL STANDARD VERIFIED 01:08
ALCO TV 29.55 in .086 01:08
BLANK TEST .000 01:09
SUBJECT SAMPLE .000 01:09
BLANK TEST .000 01:10
ALCO TV 29.55 in .088 01:11
BLANK TEST .000 01:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JANUARY 05, 2011

TIME 01:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@