

JAN 21 2011
2011
3/9/11

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Jeremy Coukling ID# 675 ✓ Date 01.17.11 ✓

A Agency VPD Phone # 835 4560

Instrument Location VPD Jail

B Alco S/N 57023 ✓ Target Value .077 ✓ High Pressure 1160

Alco Test Values .076 ✓ 1st Alco .076 ✓ 2nd Alco

Signature [Signature] 221

(OVER)

BMB
2/8/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR 2011 DLW

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JANUARY 17, 2011

OPERATOR'S NAME:
CONKLING/JEREMY/WHEELER
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: VDZ1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: 57023

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 30.24 in
ALCO TARGET .077 11:47
BLANK TEST .000 11:48
INTERNAL STANDARD VERIFIED 11:48
ALCO TV 30.24 in .076 11:48
BLANK TEST .000 11:49
SUBJECT SAMPLE .000 11:50
BLANK TEST .000 11:50
ALCO TV 30.24 in .076 11:51
BLANK TEST .000 11:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JANUARY 17, 2011
TIME 11:44

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 30.24 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~