

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


FEB 01 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130130 ✓
Name	JEREMY CONILLING ✓	ID#	6715 ✓
Date	01.28.10 ✓		
A Agency	VALDEZ PD	Phone #	835 4560
Instrument Location	VPD JAIL ✓		
B Alco S/N	70729 ✓	Target Value	.081 ✓
High Pressure	525		
Alco Test Value Average	.080 ✓	.081 ✓	
1 st Alco		2 nd Alco	
Signature	 276 ✓		0089/3/10


(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Feb, 2010.


(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130 ✓

JANUARY 28, 2010 ✓ *COB 2/3/10*

OPERATOR'S NAME:
CONKLING/JEREMY/WHEELER

OPERATOR'S NUMBER: 6715

SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC

DEPT/AGENCY: UDZ1

CASE/REPORT: UOC

TEST TYPE: U ✓

ALCO TARGET VALUE: .001 ✓

ALCO S/N: 70729

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.70 in ✓		
ALCO TARGET	.000	00:57
BLANK TEST	.000	00:58
INTERNAL STANDARD	VERIFIED	00:58
ALCO TV 29.70 in	.000 ✓	00:58
BLANK TEST	.000	00:59
SUBJECT SAMPLE <i>COB 2/3/10</i>	.000 ✓	01:00
BLANK TEST	.000	01:00
ALCO TV 29.70 in	.001 ✓	01:01
BLANK TEST	.000	01:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130 ✓

JANUARY 28, 2010 ✓
TIME 00:54

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.70 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■