

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130130 ✓		
Name	Robert Gibbs	ID#	3006	Date	4/15/08 ✓
A Agency	Valdez PD	Phone #	907 835 4560		
Instrument Location	Valdez, AK ✓				
B Alco S/N	56799 ✓	Target Value	.083	High Pressure	1100
Alco Test Value Average	.083 .080 ✓		.081 ✓		
	1 st Alco		2 nd Alco		
Signature	Robert T. Gibbs				COB 4/15/08

(OVER)

Under the Alaska Rules of Evidence, I certify that,

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert T Gibbs, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/27/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

APRIL 15, 2008

OPERATOR'S NAME:

GIBBS/ROBERT

OPERATOR'S NUMBER: 3006

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

B

O.L. #: D

DEPT/AGENCY: UDZI

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .083

ALCO S/N: 56799

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.81 in
ALCO TARGET .082 13:23
BLANK TEST .000 13:24
INTERNAL STANDARD VERIFIED 13:24
ALCO TV 29.81 in .080 13:24
BLANK TEST .000 13:25
SUBJECT SAMPLE .000 13:26
BLANK TEST .000 13:26
ALCO TV 29.81 in .081 13:27
BLANK TEST .000 13:28

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

APRIL 15, 2008
TIME 13:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~