

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 16 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130130

Name JEREMY CONKLIN ID# 6715 Date 05.07.10

A Agency VPD Phone # 835 4560

Instrument Location VPD JAIL

B Alco S/N 130130 x172988 Target Value .085 High Pressure 1000
03/24/10

Alco Test Value Average .085 .085
1st Alco 2nd Alco

Signature [Signature] 271 003
5/24/10

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of June, 2010.

[Signature]
Carolyn M. Noland
(Notary Seal Stamp)

Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

MAY 07, 2010

OPERATOR'S NAME: *CB 5/24/10*
CONKLINS/JEREMY/WHEELER
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: UDZ1
CASE/REPORT: VOC
TUBE TYPE: H
ALCO TARGET VALUE: .085
ALCO S/N: X172988

— BREATH ANALYSIS —

.085 ADJUSTED FOR 30.66 in		
ALCO TARGET	.087	08:43
BLANK TEST	.000	08:44
INTERNAL STANDARD	VERIFIED	08:44
ALCO TV 30.66 in	.085	08:44
BLANK TEST	.000	08:45
SUBJECT SAMPLE	.000	08:45
BLANK TEST	.000	08:46
ALCO TV 30.66 in	.085	08:47
BLANK TEST	.000	08:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

MAY 07, 2010
TIME 08:41

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
VENTILERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	44c
BAROMETER:	30.66 in
FLOW DETECTOR:	OKAY
PUMP	
WHEEL SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢