

NOV 09 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130130</u> ✓
Name		<u>Jeremy Conkling</u>	ID# <u>6715</u> ✓ Date <u>11.5.10</u> ✓
A	Agency	<u>UPD</u>	Phone # <u>835 4560</u>
Instrument Location <u>UPD Jail</u>			
Alco S/N		<u>57324</u> ✓	Target Value <u>.087</u> ✓ High Pressure <u>1150</u>
Alco Test Values		<u>.087</u> ✓ 1st Alco	<u>.089</u> ✓ 2nd Alco
Signature			<u>271</u>

(OVER)

BMB
12/8/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of DEC, 2010.


 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

NOVEMBER 05, 2010

BMB
12/8/10

OPERATOR'S NAME:
CONKLING/JEREMY/WHEELER
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:
VOC

SUBJECT'S FIRST NAME/MI :
VOC

D.L. #: VOC
DEPT/AGENCY: UDZ1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .087
ALCO S/N: 57324

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.51 in		
ALCO TARGET	.085	09:41
BLANK TEST	.000	09:42
INTERNAL STANDARD	VERIFIED	09:42
ALCO TU 29.51 in	.087	09:42
BLANK TEST	.000	09:43
SUBJECT SAMPLE	.000	09:44
BLANK TEST	.000	09:45
ALCO TU 29.51 in	.089	09:46
BLANK TEST	.000	09:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

NOVEMBER 05, 2010
TIME 09:30

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	29.51 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxy{|}~■