

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

MAY 18 2011

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N	<u>130131</u>
Name	<u>Shayne Calt</u>	ID#	<u>4559</u> Date <u>4-15-11</u>
A Agency	<u>AST</u>	Phone #	<u>733-2256</u>
Instrument Location <u>Talkeetna AST</u>			
Alco S/N		Target Value	High Pressure
<u>D119594</u>		<u>.078</u>	<u>500</u>
B Alco Test Value Average	<u>.077</u>	<u>.077</u>	
	1 st Alco	2 nd Alco	
Signature <u>[Signature]</u>			

(OVER)

BMB
5/18/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. NJB



[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130131

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

APRIL 15, 2011

OPERATOR'S NAME: *BMB 5/18/11*
GALT SHARNE H
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
U

O.L. #: U
DEPT/AGENCY: TAL0
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: D119594

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.94 in		
ALCO TARGET	.078	11:08
BLANK TEST	.000	11:09
INTERNAL STANDARD	VERIFIED	11:09
ALCO TV 29.94 in	.077	11:10
BLANK TEST	.000	11:11
SUBJECT SAMPLE	.000	11:11
BLANK TEST	.000	11:12
ALCO TV 29.96 in	.077	11:12
BLANK TEST	.000	11:13

BMB 5/18/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

APRIL 15, 2011
TIME 18:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-.0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~||