VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130131

Supervisor/Operator Performing the Verification Procedure:

A

Name Shayne Calt ID# 4559 Date 3-23-08

Agency AST Phone # 733-2256

Instrument Location Talkeetna AST

B

Alco S/N X124665 Target Value .079 High Pressure 700

Alco Test Value Average .077 (1st Alco) .077 (2nd Alco)

Signature [signature] CAB 3/25/08

(OVER)

(Do Not write in the area below)

Under ~~the Alaska Rules of Evidence, I certify that;~~

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Signed and sealed,

[signature] 3/26/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

MAR 25 2008
BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130131

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT [redacted]IC SAFETY
DATAMASTER cdm 130131

MARCH 23, 2008 CAB 3/25/08

OPERATOR'S NAME:
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATIONOFCAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: TAL0
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X124665

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.13 in

ALCO TARGET	.076	18:16
BLANK TEST	.000	18:17
INTERNAL STANDARD	VERIFIED	18:17
ALCO TV 29.13 in	.077	18:17
BLANK TEST	.000	18:18
SUBJECT SAMPLE	.000 CAB 3/25/08	18:19
BLANK TEST	.000	18:19
ALCO TV 29.15 in	.077	18:20
BLANK TEST	.000	18:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

MARCH 23, 2008
TIME 18:22

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.15 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~