

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 27 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130131</u> ✓
Name		<u>Shayne Calt</u>	ID# <u>4559</u> Date <u>7-16-09</u> ✓
A	Agency	<u>AST</u>	Phone # <u>733-2256</u>
Instrument Location		<u>Talkeetna</u>	✓
Alco S/N		<u>X124665</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>450</u>
B	Alco Test Value Average	<u>.080</u> ✓	<u>.079</u> ✓
		1 st Alco	2 nd Alco
Signature		<u>[Signature]</u> ✓	<u>CMB</u> <u>8/8/09</u>

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska (Scientific Crime Detection Laboratory).
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of Aug, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130131 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record

Serial Number	130131	Model	DataMaster
Location	TAL0	AST/Talkeetna	

8/13/09

Subject Information

Last Name	Verificationofcal	O.L. #	Y
First Name	Y		

Operator Information

Operator Name	Caik Shayne M	Case	
Operator #	4559	Department	TAL0

Test Results

Alco Target	.079	Alco Serial	X124665		
Nominal	0.078	Barometer	29.85	Test Type	Y
Dry Gas 1	0.080	Time	22:41:00	Test Date	07/16/2009
Subject Alcohol	0.000	Time	22:42:15	Test Time	22:39:34
Dry Gas 2	0.075	Time	22:43:23	Report	Y

8/13/09

Diagnostic Test Record

Serial Number	130131	Model	DataMaster
Location	TAL0	AST/Talkeetna	

Diagnostic Test Results

Test Date	07/16/2009	Computer	Okay
Test Time	22:37:19	Program	Okay
Software Date	02/20/2001		
Sample Chamber (C)	49		
Breath Tube (C)	43		
Barometer (in)	29.85		
Flow Detector	Okay		
High Speed Pump	Okay		
Detector	Okay		
Filter	Okay		
Quartz	Okay		
Calibration	Okay		