

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N	<u>130131</u>
Name	<u>Shayne Calt</u>	ID#	<u>4559</u> Date <u>9/18/08</u> 9-19-08 <small>COS 9/23/08</small>
A Agency	<u>AST</u>	Phone #	<u>733-2256</u>
Instrument Location <u>Talkeetna AST</u> ✓			
Alco S/N		Target Value	High Pressure
<u>X124665</u> ✓		<u>.079</u> ✓	<u>500</u>
Alco Test Value Average		Alco	
<u>.078</u> ✓		<u>.077</u> ✓	
1 st Alco		2 nd Alco	
Signature		<u>[Signature]</u> ✓ <small>COS 9/23/08</small>	

(OVER)


Under the Alaska Rules of Evidence, I certify that: (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature] 10/5/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

SEP 22 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130131 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

SEPTEMBER 18, 2000

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION/FCAL

SUBJECT'S FIRST NAME/MI:

U

U.L. #: U

DEPT/AGENCY: TAL0

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO C/H: X124665 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.38 in
ALCO TARGET .077 23:56
BLANK TEST .000 23:57
INTERNAL STANDARD VERIFIED 23:57
ALCO TU 29.38 in .078 ✓ 23:58
BLANK TEST .000 ✓ 23:59
SUBJECT SAMPLE .000 ✓ 23:59
BLANK TEST .000 00:00
ALCO TU 29.38 in .077 ✓ 00:00
BLANK TEST .000 00:01

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

SEPTEMBER 19, 2000 ✓

TIME 00:03

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.38 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy{|}~■